

CLG Microarray Test Results

Cell Line ID: EWT S8 p2 **Lab #:** CLG-49049 **Date received:** 3/09/23 **Date Reported:** 5/18/23

Contact Person: Frank Soldner **PI:** Frank Soldner **Institute:** Albert Einstein College of Medicine

Test Code: aCGH 100 **Email:** frank.soldner@einsteinmed.edu **PO #:** [REDACTED]

Mailing Address: [REDACTED]

Sample Type: Human ESC Frozen Cells **dsDNA Concentration:** 190 ng/μL **Total dsDNA:** 2.5 μg

Sex: Female **260/280 (1.7-1.9):** 1.84 **260/230 (≥1.0):** 2.33 **Array Type:** Agilent 60K Standard aCGH

Array ID Number: 252192457373_1_1 **Human Assembly:** hg38 **Reference DNA:** Agilent Euro Female

Quality Control

A sufficient amount of high quality genomic DNA, as determined by UV spec. (NanoDrop), fluorometer (Qubit) and Agarose Gel analysis, was extracted from sample EWT S8 p2 and passed our internal quality standards for aCGH labeling.

aCGH Probes (PASS/FAIL): Pass

Experimental Deviations: None

Results:

Microarray analysis did not detect any amplifications or deletions.